

0000 7809



# WASHINGTON STATE PATROL SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: 4/20/2006 INITIALS: DEC OWNER: W/SF  
 DISTRICT: 2 ADDRESS: \_\_\_\_\_  
 LOCATION: 79M Shop COUNTY: \_\_\_\_\_  
 MAKE: Hosumi TYPE: port wheel weigher ACCURACY CLASS: III  
 MODEL: 101 ID. NO: 27598 CAPACITY: 20,000 SECTIONS: 1  
 MINIMUM GRADUATIONS: 50 REG. ELEMENT: dial PLATFORM: 15" X 26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
<u>Center</u>	<u>5000</u>	<u>5000</u>	<u>0</u>				
<u>"</u>	<u>10000</u>	<u>10000</u>	<u>0</u>				
<u>"</u>	<u>15000</u>	<u>15000</u>	<u>0</u>				
<u>Center</u>	<u>20000</u>	<u>20000</u>	<u>0</u>				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cronan Date: 4/26/2006

State of Washington  
 County of Thurston  
 Signed and sworn to (or affirmed) before  
 me on 4-26-06 by

Michael J. Evans  
 (Signature of notary public)

Printed Name of Notary Michael J. EVANS  
 My appointment expires 12-27-07



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS